Approved For Release 2001/08/07 The DP81-01044R000100010016-4

CENTRAL INTELLIGENCE AGENCY INFORMATION REPORT

REPORT NO

CD NO.

COUNTRY Hungary

DATE DISTR. 6

January 51

NO. OF PAGES

PLACE ACQUIRED DATE ACQUIRED DATE OF

SUBJECT Examination of Glass Pharmaceutical Vial

TO CIA LIBRARY NO. OF ENCLS.

SUPPLEMENT TO REPORT NO.

25X1A

25X1X

one from a consignment snipped by Elektroimpex, Budapest, numgary, Albanian Ministry of Industry, during August of 1950.

- The vial was subjected to visual examination and spectro-analysis.
- Evaluation of the vial is as follows:
 - The quality and workmanship are comparable with similar USA products.
 - The spectro-analysis reveals a composition acceptable for glass ampoules by USA standards.
 - The properties of the glass closely match those of a South American, amber, ampoule glass tested by Corming several years ago.
 - The ampoule was fabricated from machine-made glassic. a; tubings to a construction to be constituted to o aj ora rose Milita .. pit. warde
 - The evaluating laboratory states that there is nothing remarkable about the ampoule or the glass from which it is made and that they are acceptable by US standards.

end =

CLASSIFICATION 3

NSRB DISTRIBUTION STATE NAVY HFC X ARMY